

Data recorded 12 December, 2009
09:30-09:50 1SPS

Upper trace Trillium Compact 120

T=120 sec
h=0.707
G=49942 V s / m
305.176 uV/count
Filter 30 sec 6-pole low pass
50 sec 2-pole high pass

Computed RMS = 2.261 nm/sec

Lower trace FBV

T=50 sec
h=0.707
G=63744 V s / m
305.176 uV/count
Filter 30 sec 6-pole low pass
120 sec 2-pole high pass

Computed RMS = 2.187 nm/sec

